

Sub: Purchase of C-Mac Video-Laryngoscope system with D-blade from M/s. Karl Storz GmbH & Co, Germany on proprietary basis

Regional Cancer centre, Thiruvananthapuram is in the process to purchase one C-Mac Video-Laryngoscope system with D-blade from M/s. Karl Storz GmbH & Co, Germany for our Anaesthesiology Dept on proprietary basis. (*Detailed technical specification and Proprietary Certificate are also attached*) This matter is being published for open information to submit objections, comments, if any from any vendor regarding the proprietary nature of the equipment. If any vendors have objection for purchasing of this equipment directly from M/s. Karl Storz on a proprietary basis, that may please be intimated to the undersigned on or before 31/08/2017 failing which, it will be presumed that any other vendor is having no objection/comment to offer and RCC will proceed the purchase.

Trivandrum
16/08/2017



DIRECTOR

ADULT FLEXIBLE INTUBATION VIDEO ENDOSCOPE & VIDEO LARYNGOSCOPE WITH SCREEN – 2
Nos

- Flexible Intubation Endoscope with CMOS chip on tip technology having NO Optical Fiber bundles. Intubation Endoscope to display Full Frame 4:3 Imaging, replacing the circular display windows.
- Screen 7 inch or more in size for display with clearly arranged soft keys for immediate use. The screen should have HDMI or HD for connecting to a big screen. Also there should be port to connect USB drive.
- Automatic/ manual white balance facility should be available on the monitor as well as on the scope
- Documentation of video & still images should be possible on data card or transferred to USB drive with JPEG and MPEG4 format which can be easily transferred to the computer/laptop. Documented videos & still images should be easily recalled on the monitor.
- Screen should be rechargeable & runs on Lithium Ion Batteries (operating time- 100 min or more)
- It should be light weight , high resolution & potable flexible scope
- Airway Guide (cum Bite block) for Oral intubation should be provided with the set (at least 10 airways)
- Integrated recording should be done on a USB drive.
- TUBE HOLDER should be a part of standard accessory
- Set should include- Suction Adaptors (Disposable), Cleaning brush & Leakage tester as standard accessories
- Suitable for following applications-
 - Bronchoscopy
 - Endotracheal Intubation (Gold standard for Difficult Airways)
 - Foreign body removal
 - Bronchial Lavage
 - Inspection of the Airways
 - Dilatation Tracheotomy

Technical Details of Flexible Video Endoscope-

- a) Tip deflection UP/DOWN: 140°/140° ,
- b) Angle of view 120° or more,
- c) Working Length: at least 65 cm,
- d) Total length: at least 93 cm,
- e) Working Channel diameter: at least 2.3 mm,
- f) Distal Tip Outer Diameter: 5.5 mm

Dr. Rupa Sreedhar

Rupa Sreedhar
 21/3/2017

[Signature]
 Ms. Shyma K. Babu

Rachel Cherian Koshy

Dr. Rachel Cherian Koshy

Video Laryngoscope should consist of following features:

- Required is Macintosh blades with closed European Metal finish D Blade with integrated camera chip and LED light illumination for obtaining more than 50000 Lux of brightness.
- It should get connected with above mention Screen 7inch or more in size for display with feature control buttons on the screen with HDMI output for connecting to a big screen.
- Magill forceps for foreign body removal and for assisting nasal intubation.
- Soft bag should be supplied to place the monitor and system can also be operated without taking monitor out from the bag.
- IV Stand for positioning the monitor , with tray for laryngoscopes should be provided
- Accessories like protection cap, tray for cleaning and sterilization of blades (at least two blades at a time) should be provided
- Blades and connection cable should be fully immersible in disinfecting solution
- One imported Trolley to hang Scope as well monitor manufactured by same company should be provided

Rupa Sreedhar
21/3/2017

Dr. Rupa Sreedhar

Shyma K. Babu
Ms. Shyma.K.Babu

Rachel Cherian Koshy

Dr. Rachel Cherian Koshy

KARL STORZ GmbH & Co. KG • P. O. Box 230 • D-78503 Tuttlingen

To whom it may concern



Your ref.

Date of your letter

Our ref.
Exp./Jun

Date
03.03.2014

Proprietary Certificate

This is to certify that M/S Karl Storz GmbH & Co. KG, Mittelstrasse 8, D-78532, Tuttlingen, Germany is the legal holders of Proprietary Rights of C-Mac brand and manufacturers of 'C-Mac Video-laryngoscope system with D- blade'.

The design and trademark of 'C-Mac Video-Laryngoscope system with D-Blade' range are exclusive rights of M/S Karl Storz GmbH & Co. KG, Tuttlingen, Germany. No individual or company is authorized to use 'C-Mac Video-Laryngoscope system with D-Blade' brand or manufacture C-Mac range.

For and on behalf of

KARL STORZ GMBH & CO.KG

i.V.


Uwe Junker

Deputy Executive Director
Global Sales Administration

Office Address:
KARL STORZ GmbH & Co. KG
Mittelstraße 8
D-78532 Tuttlingen

Bank accounts:
Volksbank Donau-Neckar eG Tuttlingen (BLZ 643 901 30)
Acc. No. 772 003, SWIFT: GENO DES1TUT
IBAN: DE97 6439 0130 0000 7720 03

Phone: +49 (0)7461 708-0
Fax: +49 (0)7461 708-105
E-Mail: info@karlstorz.de
Web: www.karlstorz.com
Telefax: DE 74465858

Commerzbank AG Tuttlingen (BLZ 643 800 11)
Acc. No. 2 713 305 00, SWIFT: COBA DE FF 643
IBAN: DE69 6438 0011 0271 3305 00

Kommanditgesellschaft:
KARL STORZ GmbH & Co. KG
Registered in Tuttlingen
Commercial Register: Stuttgart HRA 450442
VAT DE 142931059

Kreissparkasse Tuttlingen (BLZ 643 500 70)
Acc. No. 1322 SWIFT: SOLA DES 1 TUT
IBAN: DE79 6435 0070 0000 0013 22

Unlimited Partner:
KARL STORZ-Verwaltungsgesellschaft mbH
Registered in Tuttlingen
Commercial Register: Stuttgart HRB 450225
Managing Partner: Sybille Storz Dr. h. c. mult.

Deutsche Bank AG Tuttlingen (BLZ 653 700 75)
Acc. No. 211 639 000 SWIFT: DEUT DESS653
IBAN: DE09 6837 0075 0211 6390 00